



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
TTY: Contact Through Relay

Jim Doyle, Governor  
Richard J. Leinenkugel, Secretary

August 17, 2009

A-DEC INC.  
ENGINEERING  
TOM LOUISELL  
2601 CRESTVIEW DRIVE  
PO BOX 111  
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: A-DEC INC.  
Product Name: DENTAL DELIVERY SYSTEM  
Model Number(s): DECADE AUTOMATIC CONTROL: 4100, 4201, 4212, 4220, 4225 AND 4400; AND  
EXCELLENCE CONTROL 4300  
Product File No: 20090286

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of January 2015.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval supersedes the approval issued on August 19, 2004 under product file number 20040434.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20040434.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov